

## Advanced Test Equipment Rentals - www.atecorp.com 800-404-ATEC (2832)



## Stereo Zoom Microscope

## **Specifications**

Eyepiece:	Standard configuration  Working distance 100mm		Auxiliary objectives  0.5X  Working distance 165mm					
					Magnification	Field of View	Magnification	Field of View
					10X/20	7X	28.6	3.5X
	45X	4.4	22.5X	8.9				
20X/10	14X	14.3	7X	28.6				
	90X	2.2	45X	4.4				
Power Supply:	220V - 50Hz, 110V - 50/60Hz							
Transformer:	Input:	220/110VAC						
	Output:	12V DC/45W						
Illuminator:	Top light	12V/15W halogen lamp						
	Bottom light	12V/15W halogen lamp						

## **Options Include:**

- · Fluorescent ring illuminator
- Camera
- Laptop
- Scale
- Stand

Diameter of field of view (mm) = Field number of eyepiece / Zoom mag. X Auxiliary objective mag.

<sup>\*</sup> Working distance is fixed regardless of the magnification factor.

<sup>\*</sup> Total mag.= Zoom mag. X Eyepiece mag. X Auxiliary objective mag.