



Advanced Test Equipment Corp.  
www.atecorp.com 800-404-ATEC (2832)

# SciAps

---

**Inspection Site:**

**Inspection Date:** 05/31/2022 - 06/01/2022

**Instrument Type:**

**Action Level:** 1.0 (mg/cm<sup>2</sup>)

**Job Number:**

**Notes:**

---

Inspection Site:

Inspection Date: 05/31/2022 - 06/01/2022

Unit Started: 05/31/2022 03:48:15

Total Readings: 31

Unit Ended: 06/01/2022 10:34:36

Action Level: 1.0 (mg/cm<sup>2</sup>)

Classification Level: 0.5 (mg/cm<sup>2</sup>)

Test #	Result	Room	Job Number	Room Choices	Structure	Choice	Location	Substrate	Wall	Condition	Condition Choices	Lead (mg/cm <sup>2</sup> )	Mode
37	Positive	Apartment	1	Kitchen								1.16	Quick
38	Positive	Apartment	1	Kitchen	Door	Threshold	2	Wood	B	Deteriorated	Weathering	1.13	Quick
39	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.12	Quick
40	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.23	Quick
41	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.370	Quick
42	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.364	Quick
43	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.355	Quick
44	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.809	Quick
45	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.826	Quick
46	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	3.62	Quick
47	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	3.88	Quick
48	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	3.55	Quick
49	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.86	Quick
50	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.61	Quick
51	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.75	Quick
52	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.18	Quick
53	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.13	Quick
54	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.31	Quick
55	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.20	Quick
56	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.843	Quick

Inspection Site:

Inspection Date: 05/31/2022 - 06/01/2022

Unit Started: 05/31/2022 03:48:15

Total Readings: 31

Unit Ended: 06/01/2022 10:34:36

Action Level: 1.0 (mg/cm<sup>2</sup>)

Classification Level: 0.5 (mg/cm<sup>2</sup>)

Test #	Result	Room	Job Number	Room Choices	Structure	Choice	Location	Substrate	Wall	Condition	Condition Choices	Lead (mg/cm <sup>2</sup> )	Mode
57	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.823	Quick
58	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.359	Quick
59	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.358	Quick
60	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.356	Quick
61	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.370	Quick
65	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.897	Quick
66 (CAL)	PCS Pass											1.13	Timed
67	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	1.31	Timed
68	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.409	Timed
69	Negative	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.351	Quick
70	Positive	Apartment	1	Kitchen	Door	Threshold	2		B	Deteriorated	Weathering	0.781	Quick

--- End of Readings ---